



Certificate of Analysis

**Product Name: Oxford BenchMate
1.7mL Microcentrifuge Tubes Bag**
Catalog Number: OMT-017B-SL
Batch Number: EXAMPLE
Expiry Date: 05/20/2021

Customer Focus:

At Oxford Lab Products, our customers' satisfaction is exactly what we are pursuing. We have good management and product development teams who have years of industry experience. Our cell and tissue culture professionals and our entrepreneurial spirit make us responsive to customer needs. We welcome special requests and work to provide the best and most personal service in the industry.

Satisfaction Guarantee:

Oxford Lab Products, offers a total satisfaction guarantee so you can be confident in your purchasing decision. If for any reason you are not satisfied with the product performance or service provided by Oxford, we will either replace or issue a refund for the purchase price of your product.

Oxford Lab Products, are manufactured in accordance with ISO 13485: 2016 & ISO 9001: 2015 quality management systems in a class 100,000 clean room.

Sterilization - Product labeled as sterile is irradiated and do simetrically released upon ISO 11137 recommended practices in effect at the time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release. Oxford Lab Products products labeled sterile meet a minimum requirement of 10⁻⁶ SAL (Sterility Assurance Level).

Material - Free of Bisphenol A (BPA), phthalates and cytotoxic effects. Resins are USP Class-VI certified, RoHS, FDA regulation CFR 21, and are free of Substances of Very High Concern (SVHC), REACH compliant, and are FDA approved for food contact.

Non-Pyrogenic - Products labeled non-pyrogenic have been validated per FDA guidelines on LAL (Limulus Amebocyte Lysate testing for medical devices and company guidelines. The acceptance level for product is less than 0.05 EU/mL.

DNase & RNase Free - This product has been tested and is free of any detectable DNase / RNase contamination.

Trace Metal - All material are certified metal-free by material manufacturers before they are turned into production which compliance with US Pharmacopoeia guidelines.

ATP free - This product has been tested and is free of ATP.

PCR inhibitor - Products are tested via PCR amplification and gel electrophoresis analysis. Samples must show normal amplification to be considered free of PCR inhibitors.

Quality Control Testing - This product has been inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

A handwritten signature in black ink, appearing to read "Alex Alvarado".

Product Manager
Alexander Alvarado